

To: CN=Steve Vandegrift/OU=ADA/O=USEPA/C=US@EPA[]
Cc: CN=Cindy Paul/OU=ADA/O=USEPA/C=US@EPA;CN=Shauna Bennett/OU=ADA/O=USEPA/C=US@EPA;CN=Sujith Kumar/OU=ADA/O=USEPA/C=US@EPA[];
N=Shauna Bennett/OU=ADA/O=USEPA/C=US@EPA;CN=Sujith Kumar/OU=ADA/O=USEPA/C=US@EPA[]; N=Sujith Kumar/OU=ADA/O=USEPA/C=US@EPA[]
From: CN=John Cox/OU=ADA/O=USEPA/C=US
Sent: Tue 7/17/2012 12:53:57 PM
Subject: Re: Pavillion ADQ Fw: Low Molecular wt acids
[6521_6522 run log with 10 samples highlighted.pdf](#)
[RSKSOP-112 with DQO highlighted.pdf](#)
<http://www.neptuneandco.com>

Steve

Attached are two pdf's, one is the run log with the sample count highlighted and the second is Table 4. Summary of measurement quality objectives for six organic acids analysis by HPLC with the text describing our basis for the sample count. In short, we do not count spikes or duplicates in the 10 sample interval. Our question is should we do the CCC after 10 samples or 10 analyses? We would have no issues with doing a CCC every 10 analyses.

John S. Cox

Senior Scientist
Shaw Environmental, Inc
Robert S. Kerr Environmental Research Center
Ground Water and Ecosystems Restoration Division
NRML/ORD/USEPA
919 Kerr Research Drive
Ada, Oklahoma 74820
580-436-8769
cox.john@epa.gov

From: Steve Vandegrift/ADA/USEPA/US
To: Shauna Bennett/ADA/USEPA/US@EPA
Cc: John Cox/ADA/USEPA/US@EPA, Sujith Kumar/ADA/USEPA/US@EPA, Cindy Paul/ADA/USEPA/US@EPA
Date: 07/16/2012 04:57 PM
Subject: Pavillion ADQ Fw: Low Molecular wt acids

Shauna-

Please see email below.

Apparently CCCs were not run, but LCS samples were. Were these actually CCCs bu labelled as LCS?

Please provide a response tomorrow morning.

sv

Steve Vandegrift, QA Manager
Ground Water and Ecosystems Restoration Division
NRMRL/ORD/USEPA
P.O. Box 1198
919 Kerr Research Dr.
Ada, OK 74820
(580)436-8684 (voice)
(580)436-8528 (fax)
vandegrift.steve@epa.gov

----- Forwarded by Steve Vandegrift/ADA/USEPA/US on 07/16/2012 04:53 PM -----

From: David Gratson <dgratson@neptuneinc.org>
To: Steve Vandegrift/ADA/USEPA/US@EPA
Cc: Rebecca Shircliff <rshircliff@neptuneinc.org>
Date: 07/16/2012 04:17 PM
Subject: Low Molecular wt acids

Steve,

here is the low molecular weight acids issue/question:

Low MW Organic Acids QC Frequency Not Met. The continuing calibration check (CCC) frequency was not met for sample set 6521, where a CCC was not ran every 10 samples (there was a 15 field sample gap). However, there were 2 lab control spikes ran within this gap that were within control limits

See page 4, the run log in file *OA866SF SS#6521.pdf

--

David Gratson, CEAC

Environmental Chemist
Neptune and Company, Inc.
1505 15th St..Suite B
Los Alamos, NM 87544

Best way to reach me -->(505) 514-0019 (will forward to my cell phone if no answer)
Second best way ---> Cell: 505-660-8521
Work Voice: 505-662-0707 X29
Fax: 505-662-0500
<http://www.neptuneandco.com>

"Nature holds the key to our aesthetic, intellectual, cognitive and even spiritual satisfaction." E. O. Wilson.

Confidentiality Notice: This email, including all attachments is for the sole use of the intended recipient or recipients and may contain confidential and privileged information. Any unauthorized review, use, disclosure, or distribution is prohibited. If you are not the intended recipient, please contact the sender and destroy all copies of this message.